## TSK-GEL® PW<sub>XL</sub> Guardcolumn Products

Part Numbers:

08033, PW<sub>XL</sub> Guardcolumn, 6.0mm ID x 4.0cm, 12 $\mu$ m 08034, Oligo Guardcolumn , 6.0mm ID x 4.0cm, 12 $\mu$ m

This sheet contains the recommended operating conditions and the specifications for TSK-GEL PWxl and Oligo guard columns. Installation instructions and column care information are described in a separate Instruction Manual.

## A. OPERATING CONDITIONS

1. Shipping Solvent: Water

2. Max. Flow Rate: 1.0 mL/min

When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.

Standard Flow Rate: 0.5 - 0.8 mL/min

Max. Pressure: 40 kg/cm<sup>2</sup> = 600 psi

5. pH Range: 2.0 - 12.0
6. Salt Conc.: ≤ 0.5 Molar

7. Organic Conc.: ≤ 20%. It is possible to use up to 50% organic when the solvent change is made very gradually using a shallow

gradient at low flow rate.

8. Temperature: 10 - 80°C. Reduce flow rate when operating below 10°C.

9. Cleaning Solvents: (1) High salt concentration buffer (0.5 - 1.0M), or

(2) pH 2 - 3 or pH 9 - 12 buffer, or

3) Buffer with acetonitrile or methanol, or

(4) Buffer with urea or SDS

NOTE: Choose a cleaning solvent based on sample properties, e.g. use (1) to remove basic polymers, (3) to remove

hydrophobic proteins etc.

10. Storage: Store the column in a 0.05% NaN<sub>3</sub> solution or 20% ethanol in DI water when it will not be used the next day. For

overnight storage flush the column at low flow rate with the mobile phase. Prevent air from entering the column!

11. Column Protection: The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends

greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample

injections, when the peaks become excessively wide, or when the peaks show splitting.

12. TSKtop-off gel: Occasionally, due to accident, sample, mobile phase or operational variables, a depression can develop at the

guard column inlet. Use TSKtop-off gel PWXL for filling in such voids.

DS1137 Revised 30AUGUST2006

Tosoh Bioscience LLC 3604 Horizon Drive, Suite 100 King of Prussia, PA 19406 Phone: (484) 805-1219

Orders and Technical Service: (800) 366-4875

FAX: (610) 272-3028