## **TSK-GEL<sup>®</sup> G3000PW Products**

Part Numbers:	05762, 7.5mm ID x 30cm, 12μm 05106, 7.5mm ID x 60cm, 12μm 16249, 21.5mm ID x 30cm, 17μm	06762, Guard Column for P/N 05762 & 05106 06758, Guard Column for P/N 16249
---------------	---	--

This sheet contains the recommended operating conditions and the specifications for TSK-GEL G3000PW columns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATIN	G CONDITIONS	
1. Shippi	ing Solvent:	Water
2. Max. I	Flow Rate:	<ol> <li>1.2mL/min (7.5mm ID)</li> <li>8.0mL/min (21.5mm ID)</li> <li>When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow rate.</li> </ol>
3. Stand	ard Flow Rate:	0.5 - 1.0mL/min (7.5mm ID) 1.6 - 6.0mL/min (21.5mm ID)
4. Max. F	Pressure:	2MPa (7.5mm ID x 30cm) 4MPa (7.5mm ID x 60cm)
5. pH Ra	ange:	2.0 - 12.0
6. Salt C	onc.:	≤ 0.5 Molar
7. Organ	ic Conc.:	≤ 20%. It is possible to use up to 50% organic when the solvent change is made very gradually using a shallow gradient at low flow rate.
8. Temp	erature:	10 - 80°C, Reduce flow rate when operating below 10°C.
9. Clean	ing Solvents:	<ul> <li>(1) High salt concentration buffer (0.5 - 1.0M), or</li> <li>(2) pH 2 - 3 or pH 9 - 12 buffer, or</li> <li>(3) Buffer with acetonitrile or methanol, or</li> <li>(4) Buffer with urea or SDS</li> </ul>
	NOTE	Choose a cleaning solvent based on sample properties, e.g. use (1) to remove basic polymers, and (3) to remove hydrophobic proteins, etc.
10. Storaç	ge:	Store the column in a 0.05% NaN <sub>3</sub> solution or 20% ethanol in DI water when it will not be used the next day. For overnight storage flush the column at low flow rate with the mobile phase. Prevent air from entering the column!
11. Colum	nn Protection:	The use of guard columns is recommended to prolong the life of the analytical column. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced after every 30-40 sample injections, when the peaks become excessively wide, or when the peaks show splitting.
B. SPECIFICA	TIONS	
	nance of TSK-GEL G3000 rol specifications:	PW columns is tested under the conditions described in the Data Sheet. All columns have passed the following
1. Numb (N):	er of Theoretical Plates	> 5,000 (7.5mm ID x 30cm) > 10,000 (7.5mm ID x 60cm)
2. Asymi	metry Factor (AF):	0.7 - 1.6

DS1016 Revised 18AUGUST2009

**Tosoh Bioscience LLC** 3604 Horizon Drive, Suite 100 King of Prussia, PA 19406 Phone: (484) 805-1219 Orders and Technical Service: (800) 366-4875 FAX: (610) 272-3028

www.tosohbioscience.com

Tosoh Bioscience GmbH Zettachring 6 D-70567 Stuttgart Germany Phone: 0711/13257 0 FAX: 0711/13257 89